

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020425\
 Data File : VY021053.D
 Acq On : 04 Feb 2025 13:06
 Operator : SY/MD
 Sample : Q1216-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-18.1-012825

Quant Time: Feb 05 00:45:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	91498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	158532	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	140199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	53582	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	51947	55.412	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.820%
35) Dibromofluoromethane	7.634	113	54291	53.437	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.880%
50) Toluene-d8	10.109	98	192301	49.717	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.440%
62) 4-Bromofluorobenzene	12.408	95	56408	44.640	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020425\
Data File : VY021053.D
Acq On : 04 Feb 2025 13:06
Operator : SY/MD
Sample : Q1216-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-18.1-012825

Quant Time: Feb 05 00:45:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

