

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020425\
 Data File : VY021057.D
 Acq On : 04 Feb 2025 14:39
 Operator : SY/MD
 Sample : Q1232-05
 Misc : 15.19g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-46.1-012925

Quant Time: Feb 05 00:47:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	153324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	288726	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	272546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	112034	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	97642	62.156	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.320%
35) Dibromofluoromethane	7.634	113	97993	52.959	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.920%
50) Toluene-d8	10.109	98	359016	50.965	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.920%
62) 4-Bromofluorobenzene	12.407	95	115017	49.978	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020425\
Data File : VY021057.D
Acq On : 04 Feb 2025 14:39
Operator : SY/MD
Sample : Q1232-05
Misc : 15.19g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-46.1-012925

Quant Time: Feb 05 00:47:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

