

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY020520\
 Data File : VY001513.D
 Acq On : 05 Feb 2020 17:34
 Operator : SY/MD
 Sample : L1393-06
 Misc : 7.03G/5ML/MSVOA Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HAM-SB-01-1-2

Quant Time: Feb 06 07:53:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y020420S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 04 15:10:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.82	168	163715	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.71	114	305669	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	272768	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.43	152	108308	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.17	65	100712	61.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.20%	
35) Dibromofluoromethane	7.75	113	92129	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
50) Toluene-d8	10.19	98	365818	54.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.72%	
62) 4-Bromofluorobenzene	12.49	95	125929	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
Target Compounds						
16) Acetone	3.98	43	6559	13.756	ug/l	89
67) Ethyl Benzene	11.60	91	49716	5.020	ug/l	99
68) m/p-Xylenes	11.70	106	81349	21.883	ug/l	97
69) o-Xylene	12.03	106	25271	7.210	ug/l	96
95) Naphthalene	15.24	128	14868	3.024	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY020520\
Data File : VY001513.D
Acq On : 05 Feb 2020 17:34
Operator : SY/MD
Sample : L1393-06
Misc : 7.03G/5ML/MSVOA Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HAM-SB-01-1-2

Quant Time: Feb 06 07:53:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y020420S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 04 15:10:08 2020
Response via : Initial Calibration

