


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv020520\
 Data File : VY001498.D
 Acq On : 05 Feb 2020 09:34
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y020420S.M
 Title : SW846 8260
 Last Update : Tue Feb 04 15:10:08 2020



AutoFind: Scans 1762, 1763, 1764; Background Corrected with Scan 1754

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	28034	PASS
75	95	30	60	49.6	79154	PASS
95	95	100	100	100.0	159570	PASS
96	95	5	9	6.6	10530	PASS
173	174	0.00	2	0.9	971	PASS
174	95	50	100	66.9	106741	PASS
175	174	5	9	7.2	7732	PASS
176	174	95	101	96.4	102858	PASS
177	176	5	9	6.5	6662	PASS