

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
 Data File : VY021095.D
 Acq On : 05 Feb 2025 17:45
 Operator : SY/MD
 Sample : Q1281-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Feb 06 00:57:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.713	168	187827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	340704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	322006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	126233	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	118620	61.639	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.280%	
35) Dibromofluoromethane	7.634	113	113978	52.201	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.400%	
50) Toluene-d8	10.109	98	414504	49.865	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.720%	
62) 4-Bromofluorobenzene	12.408	95	144213	53.104	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 106.200%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.854	59	1312	10.061	ug/l #	75
16) Acetone	3.873	43	6428	21.419	ug/l	92
17) Carbon Disulfide	4.110	76	9336	1.662	ug/l #	90
25) 2-Butanone	6.896	43	2312	4.688	ug/l	95
29) Tetrahydrofuran	7.268	42	425	1.274	ug/l #	51
31) Cyclohexane	7.707	56	4124	1.256	ug/l #	7
36) 1,1-Dichloropropene	7.713	75	14625	4.522	ug/l #	50
37) Ethyl Acetate	6.896	43	2312	1.630	ug/l #	38
38) Carbon Tetrachloride	7.713	117	19153	5.068	ug/l #	16
51) 4-Methyl-2-Pentanone	10.000	43	1593	1.124	ug/l	89
59) 2-Hexanone	10.731	43	1462	1.602	ug/l #	23
67) Ethyl Benzene	11.524	91	295059	23.266	ug/l	99
68) m/p-Xylenes	11.627	106	410078	86.436	ug/l	97
69) o-Xylene	11.957	106	366532	82.866	ug/l	97
70) Styrene	11.957	104	22449	3.045	ug/l #	1
73) Isopropylbenzene	12.255	105	220067	22.318	ug/l	100
74) N-amyl acetate	12.054	43	6309	2.927	ug/l #	56
76) 1,2,3-Trichloropropane	12.664	75	24102	21.048	ug/l #	100
78) n-propylbenzene	12.597	91	44583	3.796	ug/l #	54
79) 2-Chlorotoluene	12.694	91	168192	25.000	ug/l #	41
80) 1,3,5-Trimethylbenzene	12.737	105	680431	85.111	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.255	75	2918	5.200	ug/l #	50
82) 4-Chlorotoluene	12.737	91	75661	11.038	ug/l #	41
83) tert-Butylbenzene	13.048	119	253439	34.877	ug/l #	55
84) 1,2,4-Trimethylbenzene	13.048	105	2042190	261.252	ug/l	99
85) sec-Butylbenzene	13.048	105	2042190	195.232	ug/l #	57
86) p-Isopropyltoluene	13.298	119	131668	15.153	ug/l #	77
89) n-Butylbenzene	13.554	91	1754437	221.326	ug/l #	43
90) Hexachloroethane	13.895	117	104312	57.545	ug/l #	17
92) 1,2-Dibromo-3-Chloropr...	14.304	75	2634	10.748	ug/l #	1
95) Naphthalene	15.145	128	19673768	5793.199	ug/l #	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
Data File : VY021095.D
Acq On : 05 Feb 2025 17:45
Operator : SY/MD
Sample : Q1281-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :

Quant Time: Feb 06 00:57:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 187827
Ion Ratio Lower Upper
168 100
99 56.8 44.2 66.4



#11
Tert butyl alcohol
Concen: 10.061 ug/l
RT: 4.854 min Scan# 514
Delta R.T. -0.018 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Tgt Ion: 59 Resp: 1312
Ion Ratio Lower Upper
59 100
57 0.0 7.4 11.0#









#16
Acetone
Concen: 21.419 ug/l
RT: 3.873 min Scan# 354
Delta R.T. -0.006 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 6428
Ion Ratio Lower Upper
43 100
58 28.7 26.6 39.8



#17
Carbon Disulfide
Concen: 1.662 ug/l
RT: 4.110 min Scan# 392
Delta R.T. -0.006 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Tgt Ion: 76 Resp: 9336
Ion Ratio Lower Upper
76 100
78 12.9 7.4 11.0#





#25

2-Butanone

Concen: 4.688 ug/l

RT: 6.896 min Scan# 849

Delta R.T. -0.006 min


Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 43 Resp: 2312

Ion Ratio Lower Upper

43 100

72 24.4 21.4 32.2



#29

Tetrahydrofuran

Concen: 1.274 ug/l

RT: 7.268 min Scan# 910

Delta R.T. 0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion: 42 Resp: 425


Ion Ratio Lower Upper

42 100

72 23.5 36.6 54.8#


71 0.0 33.8 50.6#








#31
Cyclohexane
Concen: 1.256 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 56 Resp: 4124
Ion Ratio Lower Upper
56 100
69 151.8 27.0 40.6#
84 137.5 71.8 107.6#




#33
1,2-Dichloroethane-d4
Concen: 61.639 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45



Tgt Ion: 65 Resp: 118620
Ion Ratio Lower Upper
65 100
67 52.2 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1132

Delta R.T. -0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:114 Resp: 340704

Ion Ratio Lower Upper

114 100

63 20.2 0.0 37.4

88 15.0 0.0 29.0



#35

Dibromofluoromethane

Concen: 52.201 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion:113 Resp: 113978


Ion Ratio Lower Upper

113 100

111 102.6 83.8 125.6

192 20.3 14.5 21.7









#36
1,1-Dichloropropene
Concen: 4.522 ug/l
RT: 7.713 min Scan# 911
Delta R.T. -0.128 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 14625
Ion Ratio Lower Upper
75 100
110 9.5 17.4 52.3#
77 0.0 24.7 37.1#



#37
Ethyl Acetate
Concen: 1.630 ug/l
RT: 6.896 min Scan# 849
Delta R.T. -0.091 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Tgt Ion: 43 Resp: 2312
Ion Ratio Lower Upper
43 100
61 49.4 11.0 16.6#
70 0.0 8.1 12.1#





#38

Carbon Tetrachloride

Concen: 5.068 ug/l

RT: 7.713 min Scan# 91

Delta R.T. -0.110 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:117 Resp: 19153

Ion Ratio Lower Upper

117 100

119 4.3 76.0 114.0#

121 0.0 24.2 36.2#



#50

Toluene-d8

Concen: 49.865 ug/l

RT: 10.109 min Scan# 1376

Delta R.T. -0.000 min

Lab File: VY021095.D




Acq: 05 Feb 2025 17:45


Tgt Ion: 98 Resp: 414504

Ion Ratio Lower Upper

98 100

100 64.6 52.3 78.5









#51
4-Methyl-2-Pentanone
Concen: 1.124 ug/l
RT: 10.000 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 1593
Ion Ratio Lower Upper
43 100
58 33.1 32.2 48.2



#59
2-Hexanone
Concen: 1.602 ug/l
RT: 10.731 min Scan# 1478
Delta R.T. -0.037 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Tgt Ion: 43 Resp: 1462
Ion Ratio Lower Upper
43 100
58 0.0 28.1 84.3#





#62

4-Bromofluorobenzene

Concen: 53.104 ug/l

RT: 12.408 min Scan# 1753

Delta R.T. -0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 95 Resp: 144213

Ion Ratio Lower Upper

95 100

174 83.9 0.0 160.0

176 81.9 0.0 151.8



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 1591

Delta R.T. -0.000 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion: 117 Resp: 322006


Ion Ratio Lower Upper

117 100

82 56.4 43.8 65.8

119 32.8 26.5 39.7





#67

Ethyl Benzene

Concen: 23.266 ug/l

RT: 11.524 min Scan# 1608

Delta R.T. -0.000 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 295059

Ion Ratio Lower Upper

91 100

106 30.8 25.0 37.4



#68

m/p-Xylenes

Concen: 86.436 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.006 min

Lab File: VY021095.D




Acq: 05 Feb 2025 17:45


Tgt Ion:106 Resp: 410078

Ion Ratio Lower Upper

106 100

91 205.4 160.7 241.1





#69

o-Xylene

Concen: 82.866 ug/l

RT: 11.957 min Scan# 1

Delta R.T. -0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:106 Resp: 366532

Ion Ratio Lower Upper

106 100

91 217.4 106.5 319.5



#70

Styrene

Concen: 3.045 ug/l

RT: 11.957 min Scan# 1679

Delta R.T. -0.018 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion:104 Resp: 22449


Ion Ratio Lower Upper

104 100

78 268.1 39.2 58.8#

103 234.9 43.8 65.6#









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.353 min Scan# 1908
Delta R.T. -0.000 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:152 Resp: 126233
Ion Ratio Lower Upper
152 100
115 0.0 30.0 90.0#
150 155.3 0.0 344.6



#73
Isopropylbenzene
Concen: 22.318 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. -0.006 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45


Tgt Ion:105 Resp: 220067
Ion Ratio Lower Upper
105 100
120 25.9 13.0 39.0








#74
N-amy1 acetate
Concen: 2.927 ug/l
RT: 12.054 min Scan# 1795
Delta R.T. -0.018 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43 Resp: 6309
Ion Ratio Lower Upper
43 100
70 14.5 36.6 55.0#
55 10.5 22.2 33.4#
61 0.0 20.2 30.4#



#76
1,2,3-Trichloropropane
Concen: 21.048 ug/l
RT: 12.664 min Scan# 1795
Delta R.T. 0.104 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Tgt Ion: 75 Resp: 24102
Ion Ratio Lower Upper
75 100
77 1374.1 0.0 0.0#





#78

n-propylbenzene

Concen: 3.796 ug/l

RT: 12.597 min Scan# 1

Delta R.T. -0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 44583

Ion Ratio Lower Upper

91 100

120 0.0 11.0 33.0#



#79

2-Chlorotoluene

Concen: 25.000 ug/l

RT: 12.694 min Scan# 1800

Delta R.T. 0.006 min

Lab File: VY021095.D




Acq: 05 Feb 2025 17:45


Tgt Ion: 91 Resp: 168192

Ion Ratio Lower Upper

91 100

126 0.0 17.0 51.0#





#80

1,3,5-Trimethylbenzene

Concen: 85.111 ug/l

RT: 12.737 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:105 Resp: 680431

Ion Ratio Lower Upper

105 100

120 48.8 24.3 72.8



#81

trans-1,4-Dichloro-2-butene

Concen: 5.200 ug/l

RT: 12.255 min Scan# 1728

Delta R.T. -0.055 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion: 75 Resp: 2918


Ion Ratio Lower Upper

75 100

53 119.8 74.0 111.0#

89 106.6 36.5 54.7#





#82

4-Chlorotoluene

Concen: 11.038 ug/l

RT: 12.737 min Scan# 1807

Delta R.T. -0.049 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 75661

Ion Ratio Lower Upper

91 100

126 0.5 17.1 51.2#



#83

tert-Butylbenzene

Concen: 34.877 ug/l

RT: 13.048 min Scan# 1858

Delta R.T. 0.043 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion: 119 Resp: 253439


Ion Ratio Lower Upper

119 100

91 95.4 31.5 94.5#

134 0.0 13.1 39.3#





#84

1,2,4-Trimethylbenzene

Concen: 261.252 ug/l

RT: 13.048 min Scan# 1858

Delta R.T. -0.000 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:105 Resp: 2042190

Ion Ratio Lower Upper

105 100

120 45.1 22.1 66.3



#85

sec-Butylbenzene

Concen: 195.232 ug/l

RT: 13.048 min Scan# 1858

Delta R.T. -0.134 min

Lab File: VY021095.D




Acq: 05 Feb 2025 17:45


Tgt Ion:105 Resp: 2042190

Ion Ratio Lower Upper

105 100

134 0.0 9.7 28.9#









#86
p-Isopropyltoluene
Concen: 15.153 ug/l
RT: 13.298 min Scan# 1191
Delta R.T. -0.000 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:119 Resp: 131668
Ion Ratio Lower Upper
119 100
134 26.1 13.1 39.1
91 0.0 11.7 35.0#



#89
n-Butylbenzene
Concen: 221.326 ug/l
RT: 13.554 min Scan# 1941
Delta R.T. -0.067 min
Lab File: VY021095.D
Acq: 05 Feb 2025 17:45

Tgt Ion: 91 Resp: 1754437
Ion Ratio Lower Upper
91 100
92 10.1 26.4 79.2#
134 0.0 12.3 36.8#





#90

Hexachloroethane

Concen: 57.545 ug/l

RT: 13.895 min Scan# 1996

Delta R.T. 0.006 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:117 Resp: 104312

Ion Ratio Lower Upper

117 100

201 0.0 33.4 100.1#



#92

1,2-Dibromo-3-Chloropropane

Concen: 10.748 ug/l

RT: 14.304 min Scan# 2064

Delta R.T. 0.024 min

Lab File: VY021095.D

Acq: 05 Feb 2025 17:45




Tgt Ion: 75 Resp: 2634


Ion Ratio Lower Upper

75 100

155 0.0 42.1 126.3#


157 0.0 55.8 167.4#





#95
 Naphthalene
 Concen: 5793.199 ug/l
 RT: 15.145 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: VY021095.D
 Acq: 05 Feb 2025 17:45

Instrument :
 MSVOA_Y
 ClientSampleId :



Tgt Ion:128 Resp:19673768
 Ion Ratio Lower Upper
 128 100
 127 20.7 10.2 15.4#
 129 18.3 8.6 12.8#

