

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
 Data File : VY021075.D
 Acq On : 05 Feb 2025 09:51
 Operator : SY/MD
 Sample : VY0205SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0205SBL01

Quant Time: Feb 06 00:24:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	173976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	323562	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	290166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	107232	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	101161	56.752	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.500%
35) Dibromofluoromethane	7.646	113	105621	50.936	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.880%
50) Toluene-d8	10.115	98	393322	49.823	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.640%
62) 4-Bromofluorobenzene	12.414	95	115413	44.750	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY020525\
Data File : VY021075.D
Acq On : 05 Feb 2025 09:51
Operator : SY/MD
Sample : VY0205SBL01
Misc : 5.00g/5.0mL/MSV0A_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
VY0205SBL01

Quant Time: Feb 06 00:24:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

