

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
 Data File : VY021087.D
 Acq On : 05 Feb 2025 14:38
 Operator : SY/MD
 Sample : Q1289-02
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FL-DRUMS-A

Quant Time: Feb 06 00:35:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	6211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	6836	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	2259	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	395	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	1552	24.388	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	48.780%#
35) Dibromofluoromethane	7.640	113	2301	52.523	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.040%
50) Toluene-d8	10.109	98	5963	35.752	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	71.500%
62) 4-Bromofluorobenzene	12.408	95	436	8.002	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	16.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
Data File : VY021087.D
Acq On : 05 Feb 2025 14:38
Operator : SY/MD
Sample : Q1289-02
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FL-DRUMS-A

Quant Time: Feb 06 00:35:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

