

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
 Data File : VY021093.D
 Acq On : 05 Feb 2025 16:58
 Operator : SY/MD
 Sample : Q1286-02
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-020425-VOC

Quant Time: Feb 06 00:56:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	203936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	340354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	307078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	127720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	102920	49.256	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.520%
35) Dibromofluoromethane	7.640	113	107906	49.471	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.940%
50) Toluene-d8	10.109	98	410838	49.475	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.940%
62) 4-Bromofluorobenzene	12.407	95	130136	47.970	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.940%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.047	105	13754	1.739	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
Data File : VY021093.D
Acq On : 05 Feb 2025 16:58
Operator : SY/MD
Sample : Q1286-02
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-020425-VOC

Quant Time: Feb 06 00:56:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

