



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020525\
Data File : VY021073.D
Acq On : 05 Feb 2025 08:16
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Title : SW846 8260
Last Update : Mon Feb 03 13:08:38 2025



AutoFind: Scans 1754, 1755, 1756; Background Corrected with Scan 1746

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	14227	PASS
75	95	30	60	56.0	36253	PASS
95	95	100	100	100.0	64752	PASS
96	95	5	9	6.6	4285	PASS
173	174	0.00	2	1.7	941	PASS
174	95	50	100	85.2	55189	PASS
175	174	5	9	7.8	4301	PASS
176	174	95	101	97.6	53875	PASS
177	176	5	9	6.6	3555	PASS