

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020623\
 Data File : VY012505.D
 Acq On : 06 Feb 2023 19:06
 Operator : KP/MD
 Sample : 01414-06RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SASP2-123-2(75-80)RE

Quant Time: Feb 07 03:47:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 07 03:42:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	158911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	268093	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	265890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	126603	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	93425	47.377	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.760%
35) Dibromofluoromethane	7.722	113	78584	41.557	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.120%
50) Toluene-d8	10.179	98	321573	40.911	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	81.820%
62) 4-Bromofluorobenzene	12.483	95	108292	40.993	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020623\
Data File : VY012505.D
Acq On : 06 Feb 2023 19:06
Operator : KP/MD
Sample : 01414-06RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SASP2-123-2(75-80)RE

Quant Time: Feb 07 03:47:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020623S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 07 03:42:26 2023
Response via : Initial Calibration

