

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020624\
 Data File : VY017254.D
 Acq On : 06 Feb 2024 10:51
 Operator : SY/MD
 Sample : VY0206SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0206SBL01

Quant Time: Feb 07 00:05:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	76909	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.710	114	143363	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	130424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	48060	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	46097	67.093	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.180%
35) Dibromofluoromethane	7.734	113	49535	63.993	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	127.980%
50) Toluene-d8	10.197	98	169018	59.971	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	119.940%
62) 4-Bromofluorobenzene	12.495	95	47300	51.233	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020624\
Data File : VY017254.D
Acq On : 06 Feb 2024 10:51
Operator : SY/MD
Sample : VY0206SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0206SBL01

Quant Time: Feb 07 00:05:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

