

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020624\
 Data File : VY017257.D
 Acq On : 06 Feb 2024 12:18
 Operator : SY/MD
 Sample : P1334-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-01-020124

Quant Time: Feb 06 12:36:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	77386	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	144347	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	124881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	39820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	44097	63.787	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.580%
35) Dibromofluoromethane	7.734	113	49275	63.223	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	126.440%
50) Toluene-d8	10.197	98	169494	59.730	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	119.460%
62) 4-Bromofluorobenzene	12.501	95	43892	47.218	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.440%
Target Compounds						
20) Methylene Chloride	4.740	84	5310	3.094	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020624\
Data File : VY017257.D
Acq On : 06 Feb 2024 12:18
Operator : SY/MD
Sample : P1334-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-020124

Quant Time: Feb 06 12:36:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

