

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020624\
 Data File : VY017269.D
 Acq On : 06 Feb 2024 17:00
 Operator : SY/MD
 Sample : P1334-05RE
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-03-020124RE

Quant Time: Feb 06 17:16:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	72695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	133427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	117563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	43501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	50071	77.102	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.200%
35) Dibromofluoromethane	7.728	113	45706	63.443	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	126.880%
50) Toluene-d8	10.185	98	155604	59.323	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	118.640%
62) 4-Bromofluorobenzene	12.489	95	43672	50.826	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.660%
Target Compounds						
20) Methylene Chloride	4.716	84	5343	3.582	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020624\
Data File : VY017269.D
Acq On : 06 Feb 2024 17:00
Operator : SY/MD
Sample : P1334-05RE
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-03-020124RE

Quant Time: Feb 06 17:16:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

