

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020625\
 Data File : VY021113.D
 Acq On : 06 Feb 2025 15:58
 Operator : SY/MD
 Sample : Q1287-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BAY-DRUMS-A

Quant Time: Feb 07 02:33:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	3112	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	4819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	3112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	1992	62.475	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.940%	
35) Dibromofluoromethane	7.646	113	1930	62.493	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.980%	
50) Toluene-d8	10.115	98	5426	46.149	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 92.300%	
62) 4-Bromofluorobenzene	12.414	95	583	15.178	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 30.360%	
Target Compounds						
86) p-Isopropyltoluene	13.298	119	1033	36.337	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020625\
Data File : VY021113.D
Acq On : 06 Feb 2025 15:58
Operator : SY/MD
Sample : Q1287-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BAY-DRUMS-A

Quant Time: Feb 07 02:33:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

