

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020625\
 Data File : VY021125.D
 Acq On : 06 Feb 2025 20:39
 Operator : SY/MD
 Sample : Q1299-03
 Misc : 4.48g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1-VOC

Quant Time: Feb 07 02:38:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	149197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	270546	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	246769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	97937	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	85553	55.967	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.940%
35) Dibromofluoromethane	7.640	113	89399	51.561	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.120%
50) Toluene-d8	10.109	98	323833	49.059	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.120%
62) 4-Bromofluorobenzene	12.414	95	103787	48.128	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.260%
Target Compounds						
95) Naphthalene	15.151	128	4448	1.688	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020625\
Data File : VY021125.D
Acq On : 06 Feb 2025 20:39
Operator : SY/MD
Sample : Q1299-03
Misc : 4.48g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1-VOC

Quant Time: Feb 07 02:38:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

