

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020625\
 Data File : VY021126.D
 Acq On : 06 Feb 2025 21:02
 Operator : SY/MD
 Sample : Q1299-07
 Misc : 5.32g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2-VOC

Quant Time: Feb 07 02:38:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	227448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	423672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	399507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	168588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	138312	59.351	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.700%
35) Dibromofluoromethane	7.640	113	140367	51.697	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.400%
50) Toluene-d8	10.109	98	516998	50.015	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.040%
62) 4-Bromofluorobenzene	12.408	95	179242	53.077	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.160%
Target Compounds						
16) Acetone	3.879	43	5210	14.336	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020625\
Data File : VY021126.D
Acq On : 06 Feb 2025 21:02
Operator : SY/MD
Sample : Q1299-07
Misc : 5.32g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2-VOC

Quant Time: Feb 07 02:38:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

