

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020723\
 Data File : VY012527.D
 Acq On : 07 Feb 2023 18:34
 Operator : KP/MD
 Sample : 01465-05
 Misc : 2.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOT-VOC-GRAB-4

Quant Time: Feb 08 01:48:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	170	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	144	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	187	86.504	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	173.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.185	98	367	35.610	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	71.220%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020723\
Data File : VY012527.D
Acq On : 07 Feb 2023 18:34
Operator : KP/MD
Sample : 01465-05
Misc : 2.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOT-VOC-GRAB-4

Quant Time: Feb 08 01:48:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

