

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020723\
 Data File : VY012524.D
 Acq On : 07 Feb 2023 17:24
 Operator : KP/MD
 Sample : 01445-01
 Misc : 3.19g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-26-020323

Quant Time: Feb 08 01:47:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	138321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	232282	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	216469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85932	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	87157	49.551	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.100%
35) Dibromofluoromethane	7.716	113	67588	39.955	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.900%
50) Toluene-d8	10.179	98	273074	38.783	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.560%
62) 4-Bromofluorobenzene	12.483	95	80935	32.955	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	65.920%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020723\
Data File : VY012524.D
Acq On : 07 Feb 2023 17:24
Operator : KP/MD
Sample : 01445-01
Misc : 3.19g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-26-020323

Quant Time: Feb 08 01:47:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY012524.D
Acq: 07 Feb 2023 17:24




Instrument :
MSVOA_Y
ClientSampleId :
JPP-26-020323


Tgt Ion:168 Resp: 138321
Ion Ratio Lower Upper
168 100
99 53.5 40.8 61.2



#33
1,2-Dichloroethane-d4
Concen: 49.551 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. 0.000 min
Lab File: VY012524.D
Acq: 07 Feb 2023 17:24

Tgt Ion: 65 Resp: 87157
Ion Ratio Lower Upper
65 100
67 48.6 0.0 98.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012524.D

Acq: 07 Feb 2023 17:24

Instrument :

MSVOA_Y

ClientSampleId :

JPP-26-020323





Tgt Ion:114 Resp: 232282

Ion Ratio Lower Upper

114 100

63 16.3 0.0 33.6

88 13.9 0.0 29.8



#35

Dibromofluoromethane

Concen: 39.955 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY012524.D

Acq: 07 Feb 2023 17:24




Tgt Ion:113 Resp: 67588


Ion Ratio Lower Upper

113 100

111 102.8 81.9 122.9

192 21.6 17.5 26.3





#50

Toluene-d8

Concen: 38.783 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. 0.000 min

Lab File: VY012524.D

Acq: 07 Feb 2023 17:24

Instrument :

MSVOA_Y

ClientSampleId :





JPP-26-020323

Tgt Ion: 98 Resp: 273074

Ion Ratio Lower Upper

98 100

100 65.2 52.2 78.2



#62

4-Bromofluorobenzene

Concen: 32.955 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VY012524.D

Acq: 07 Feb 2023 17:24




Tgt Ion: 95 Resp: 80935


Ion Ratio Lower Upper

95 100

174 97.6 0.0 188.0

176 94.4 0.0 183.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY012524.D
Acq: 07 Feb 2023 17:24

Instrument :
MSVOA_Y
ClientSampleId :
JPP-26-020323

Tgt Ion:117 Resp: 216469
Ion Ratio Lower Upper
117 100
82 51.0 41.7 62.5
119 31.9 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY012524.D
Acq: 07 Feb 2023 17:24

Tgt Ion:152 Resp: 85932
Ion Ratio Lower Upper
152 100
115 56.5 28.1 84.3
150 158.2 0.0 351.0

