

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020724\
 Data File : VY017280.D
 Acq On : 07 Feb 2024 11:13
 Operator : SY/MD
 Sample : VY0207SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0207SBL01

Quant Time: Feb 08 01:40:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	88210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	166029	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	139813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	53889	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	49432	62.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.460%
35) Dibromofluoromethane	7.722	113	48355	53.940	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.880%
50) Toluene-d8	10.185	98	169415	51.906	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.483	95	48717	45.564	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020724\
Data File : VY017280.D
Acq On : 07 Feb 2024 11:13
Operator : SY/MD
Sample : VY0207SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0207SBL01

Quant Time: Feb 08 01:40:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

