

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020724\
 Data File : VY017297.D
 Acq On : 07 Feb 2024 18:24
 Operator : SY/MD
 Sample : P1391-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-1

Quant Time: Feb 08 01:46:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	69904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	138275	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	126234	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.428	152	49852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	37451	59.972	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.940%
35) Dibromofluoromethane	7.716	113	34418	46.100	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.200%
50) Toluene-d8	10.179	98	125156	46.042	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.080%
62) 4-Bromofluorobenzene	12.483	95	38269	42.976	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.960%
Target Compounds						
52) Toluene	10.246	92	2277	1.026	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020724\
Data File : VY017297.D
Acq On : 07 Feb 2024 18:24
Operator : SY/MD
Sample : P1391-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-1

Quant Time: Feb 08 01:46:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

