



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021023\
Data File : VY012569.D
Acq On : 10 Feb 2023 09:03
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Title : SW846 8260
Last Update : Wed Feb 08 01:09:03 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.8	11673	PASS
75	95	30	60	50.4	35000	PASS
95	95	100	100	100.0	69443	PASS
96	95	5	9	6.4	4476	PASS
173	174	0.00	2	1.2	808	PASS
174	95	50	100	97.0	67368	PASS
175	174	5	9	7.7	5188	PASS
176	174	95	101	97.5	65677	PASS
177	176	5	9	6.9	4547	PASS