

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021121\
 Data File : VY003997.D
 Acq On : 11 Feb 2021 12:17
 Operator : SY/MD
 Sample : VY0211SBL01
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0211SBL01

Quant Time: Feb 11 13:21:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	138053	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	232018	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	217561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	103231	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	84278	55.52	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.04%
35) Dibromofluoromethane	7.722	113	69537	51.26	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.52%
50) Toluene-d8	10.179	98	286558	53.83	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.66%
62) 4-Bromofluorobenzene	12.477	95	94416	48.65	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.30%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021121\
Data File : VY003997.D
Acq On : 11 Feb 2021 12:17
Operator : SY/MD
Sample : VY0211SBL01
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0211SBL01

Quant Time: Feb 11 13:21:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

