

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021220\
 Data File : VY001608.D
 Acq On : 12 Feb 2020 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 13 07:53:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y020420S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 04 15:10:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.489	0.425	13.1	69	0.00
3 P	Chloromethane	0.512	0.560	-9.4	90	0.00
4 C	Vinyl Chloride	0.531	0.587	-10.5#	89	0.00
5 T	Bromomethane	0.334	0.376	-12.6	90	0.00
6 T	Chloroethane	0.320	0.371	-15.9	93	0.00
7 T	Trichlorofluoromethane	0.879	0.762	13.3	71	0.00
8 T	Diethyl Ether	0.297	0.297	0.0	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.529	0.489	7.6	75	0.00
10 T	Methyl Iodide	0.674	0.618	8.3	71	0.00
11 T	Tert butyl alcohol	0.077	0.052	32.5#	84	0.00
12 CM	1,1-Dichloroethene	0.526	0.492	6.5#	76	0.00
13 T	Acrolein	0.048	0.048	0.0	89	0.00
14 T	Allyl chloride	0.804	0.867	-7.8	88	0.00
15 T	Acrylonitrile	0.136	0.156	-14.7	99	0.00
16 T	Acetone	0.146	0.171	-17.1	98	0.00
17 T	Carbon Disulfide	1.673	1.564	6.5	75	0.00
18 T	Methyl Acetate	0.311	0.352	-13.2	101	0.00
19 T	Methyl tert-butyl Ether	1.398	1.362	2.6	82	0.01
20 T	Methylene Chloride	0.655	0.562	14.2	78	0.01
21 T	trans-1,2-Dichloroethene	0.584	0.565	3.3	79	0.00
22 T	Diisopropyl ether	1.593	1.823	-14.4	94	0.00
23 T	Vinyl Acetate	1.014	1.205	-18.8	98	0.00
24 P	1,1-Dichloroethane	0.958	0.971	-1.4	83	0.00
25 T	2-Butanone	0.189	0.233	-23.3	105	0.00
26 T	2,2-Dichloropropane	0.917	0.848	7.5	78	0.00
27 T	cis-1,2-Dichloroethene	0.638	0.621	2.7	80	0.00
28 T	Bromochloromethane	0.326	0.435	-33.4#	99	0.00
29 T	Tetrahydrofuran	0.113	0.139	-23.0	106	0.00
30 C	Chloroform	0.971	0.933	3.9#	78	0.00
31 T	Cyclohexane	0.980	0.990	-1.0	85	0.00
32 T	1,1,1-Trichloroethane	0.880	0.813	7.6	75	0.00
33 S	1,2-Dichloroethane-d4	0.503	0.530	-5.4	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.288	0.289	-0.3	86	0.00
36 T	1,1-Dichloropropene	0.457	0.451	1.3	80	0.00
37 T	Ethyl Acetate	0.221	0.265	-19.9	104	0.00
38 T	Carbon Tetrachloride	0.443	0.404	8.8	73	0.00
39 T	Methylcyclohexane	0.602	0.584	3.0	78	0.00
40 TM	Benzene	1.305	1.322	-1.3	83	0.00
41 T	Methacrylonitrile	0.103	0.097	5.8	70	0.00
42 TM	1,2-Dichloroethane	0.363	0.355	2.2	82	0.00
43 T	Isopropyl Acetate	0.438	0.518	-18.3	102	0.00
44 TM	Trichloroethene	0.352	0.332	5.7	77	0.00
45 C	1,2-Dichloropropane	0.315	0.334	-6.0#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021220\
 Data File : VY001608.D
 Acq On : 12 Feb 2020 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 13 07:53:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y020420S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 04 15:10:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.177	0.174	1.7	82	0.00
47 T	Bromodichloromethane	0.431	0.424	1.6	80	0.00
48 T	Methyl methacrylate	0.199	0.234	-17.6	100	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	86	0.00
50 S	Toluene-d8	1.091	1.198	-9.8	88	0.00
51 T	4-Methyl-2-Pentanone	0.216	0.269	-24.5	108	0.00
52 CM	Toluene	0.828	0.835	-0.8#	83	0.00
53 T	t-1,3-Dichloropropene	0.469	0.467	0.4	82	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.536	0.6	81	0.00
55 T	1,1,2-Trichloroethane	0.254	0.260	-2.4	85	0.00
56 T	Ethyl methacrylate	0.361	0.394	-9.1	91	0.00
57 T	1,3-Dichloropropane	0.447	0.457	-2.2	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.133	0.156	-17.3	98	0.00
59 T	2-Hexanone	0.157	0.203	-29.3#	109	0.00
60 T	Dibromochloromethane	0.299	0.290	3.0	79	0.00
61 T	1,2-Dibromoethane	0.250	0.248	0.8	83	0.00
62 S	4-Bromofluorobenzene	0.430	0.446	-3.7	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.328	0.297	9.5	77	0.00
65 PM	Chlorobenzene	0.982	0.949	3.4	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.344	0.327	4.9	80	0.00
67 C	Ethyl Benzene	1.815	1.769	2.5#	82	0.00
68 T	m/p-Xylenes	0.681	0.669	1.8	83	0.00
69 T	o-Xylene	0.643	0.636	1.1	82	0.00
70 T	Styrene	1.111	1.109	0.2	84	0.00
71 P	Bromoform	0.191	0.191	0.0	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.813	3.552	6.8	82	0.00
74 T	N-amyl acetate	0.991	1.135	-14.5	105	0.00
75 P	1,1,2,2-Tetrachloroethane	0.725	0.762	-5.1	96	0.00
76 T	1,2,3-Trichloropropane	0.538	0.527	2.0	90	0.00
77 T	Bromobenzene	0.825	0.755	8.5	81	0.00
78 T	n-propylbenzene	4.565	4.379	4.1	85	0.00
79 T	2-Chlorotoluene	2.602	2.416	7.1	83	0.00
80 T	1,3,5-Trimethylbenzene	3.212	2.961	7.8	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.278	0.276	0.7	91	0.00
82 T	4-Chlorotoluene	2.728	2.553	6.4	83	0.00
83 T	tert-Butylbenzene	2.715	2.563	5.6	83	0.00
84 T	1,2,4-Trimethylbenzene	3.179	2.967	6.7	83	0.00
85 T	sec-Butylbenzene	3.817	3.685	3.5	85	0.00
86 T	p-Isopropyltoluene	3.363	3.215	4.4	84	0.00
87 T	1,3-Dichlorobenzene	1.582	1.502	5.1	86	0.00
88 T	1,4-Dichlorobenzene	1.584	1.514	4.4	86	0.00
89 T	n-Butylbenzene	3.319	3.296	0.7	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021220\
Data File : VY001608.D
Acq On : 12 Feb 2020 11:15
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 13 07:53:12 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y020420S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 04 15:10:08 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.646	0.618	4.3	84	0.00
91 T	1,2-Dichlorobenzene	1.433	1.375	4.0	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.134	0.122	9.0	87	0.00
93 T	1,2,4-Trichlorobenzene	0.966	0.885	8.4	83	0.00
94 T	Hexachlorobutadiene	0.490	0.452	7.8	83	0.00
95 T	Naphthalene	2.270	2.087	8.1	89	0.00
96 T	1,2,3-Trichlorobenzene	0.853	0.789	7.5	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6