

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021221\
 Data File : VY004021.D
 Acq On : 12 Feb 2021 13:37
 Operator : SY/MD
 Sample : VY0212SBL01
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0212SBL01

Quant Time: Feb 13 00:10:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	86974	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	143640	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	131316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	63608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	54325	56.80	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.60%	
35) Dibromofluoromethane	7.722	113	44836	53.39	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.78%	
50) Toluene-d8	10.179	98	176395	53.52	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.04%	
62) 4-Bromofluorobenzene	12.477	95	62170	51.75	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.50%	
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	6740	6.76	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021221\
Data File : VY004021.D
Acq On : 12 Feb 2021 13:37
Operator : SY/MD
Sample : VY0212SBL01
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0212SBL01

Quant Time: Feb 13 00:10:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

