

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021221\
 Data File : VY004024.D
 Acq On : 12 Feb 2021 15:37
 Operator : SY/MD
 Sample : M1403-02
 Misc : 6.69g/5mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11939

Quant Time: Feb 13 00:12:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	82466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	134069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	121982	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	57382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	50324	55.50	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.00%
35) Dibromofluoromethane	7.722	113	42046	53.64	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.28%
50) Toluene-d8	10.179	98	163093	53.02	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.04%
62) 4-Bromofluorobenzene	12.477	95	56385	50.28	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.56%
Target Compounds						
16) Acetone	3.960	43	1945	8.74	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021221\
Data File : VY004024.D
Acq On : 12 Feb 2021 15:37
Operator : SY/MD
Sample : M1403-02
Misc : 6.69g/5mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11939

Quant Time: Feb 13 00:12:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

