

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021221\
 Data File : VY004025.D
 Acq On : 12 Feb 2021 16:00
 Operator : SY/MD
 Sample : M1395-02
 Misc : 5.81g/5mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FK-BPM-03-020-ABCD

Quant Time: Feb 13 00:12:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	82297	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	135007	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	127168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	62220	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	52124	57.60	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.20%
35) Dibromofluoromethane	7.722	113	43649	55.30	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.60%
50) Toluene-d8	10.179	98	166428	53.73	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.46%
62) 4-Bromofluorobenzene	12.477	95	59804	52.96	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021221\
Data File : VY004025.D
Acq On : 12 Feb 2021 16:00
Operator : SY/MD
Sample : M1395-02
Misc : 5.81g/5mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FK-BPM-03-020-ABCD

Quant Time: Feb 13 00:12:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

