

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021224\
 Data File : VY017352.D
 Acq On : 12 Feb 2024 14:40
 Operator : SY/MD
 Sample : VY0212SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0212SBL01

Quant Time: Feb 12 15:05:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	91905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	170475	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	141592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	55552	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	43683	53.206	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.420%
35) Dibromofluoromethane	7.728	113	45489	49.420	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.840%
50) Toluene-d8	10.191	98	159074	47.466	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.489	95	44841	40.845	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021224\
Data File : VY017352.D
Acq On : 12 Feb 2024 14:40
Operator : SY/MD
Sample : VY0212SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0212SBL01

Quant Time: Feb 12 15:05:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

