

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021224\
 Data File : VY017359.D
 Acq On : 12 Feb 2024 17:29
 Operator : SY/MD
 Sample : P1448-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-020924RE

Quant Time: Feb 13 01:29:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	28096	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	50159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	42296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	14108	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	13382	53.316	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.640%
35) Dibromofluoromethane	7.722	113	13682	50.519	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.040%
50) Toluene-d8	10.185	98	43612	44.229	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.460%
62) 4-Bromofluorobenzene	12.489	95	11065	34.255	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021224\
Data File : VY017359.D
Acq On : 12 Feb 2024 17:29
Operator : SY/MD
Sample : P1448-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-020924RE

Quant Time: Feb 13 01:29:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

