

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021325\
 Data File : VY021196.D
 Acq On : 13 Feb 2025 17:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 00:46:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	63	0.00
2 T	Dichlorodifluoromethane	0.446	0.457	-2.5	65	0.00
3 P	Chloromethane	0.440	0.445	-1.1	67	0.00
4 C	Vinyl Chloride	0.443	0.460	-3.8#	68	0.00
5 T	Bromomethane	0.274	0.268	2.2	66	0.00
6 T	Chloroethane	0.263	0.281	-6.8	70	0.00
7 T	Trichlorofluoromethane	0.806	0.853	-5.8	68	0.00
8 T	Diethyl Ether	0.252	0.262	-4.0	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.506	0.544	-7.5	69	0.00
10 T	Methyl Iodide	0.562	0.572	-1.8	62	0.00
11 T	Tert butyl alcohol	0.035	0.037	-5.7	74	-0.02
12 CM	1,1-Dichloroethene	0.477	0.496	-4.0#	67	0.00
13 T	Acrolein	0.053	0.032	39.6#	39#	0.00
14 T	Allyl chloride	0.803	0.852	-6.1	67	0.00
15 T	Acrylonitrile	0.105	0.120	-14.3	71	0.00
16 T	Acetone	0.080	0.091	-13.7	74	0.00
17 T	Carbon Disulfide	1.495	1.473	1.5	63	0.00
18 T	Methyl Acetate	0.280	0.336	-20.0	74	-0.01
19 T	Methyl tert-butyl Ether	1.205	1.337	-11.0	69	0.00
20 T	Methylene Chloride	0.497	0.524	-5.4	68	0.00
21 T	trans-1,2-Dichloroethene	0.525	0.544	-3.6	66	0.00
22 T	Diisopropyl ether	1.656	1.871	-13.0	71	0.00
23 T	Vinyl Acetate	0.923	1.056	-14.4	70	0.00
24 P	1,1-Dichloroethane	0.970	1.056	-8.9	70	0.00
25 T	2-Butanone	0.131	0.154	-17.6	72	-0.01
26 T	2,2-Dichloropropane	0.898	0.966	-7.6	70	0.00
27 T	cis-1,2-Dichloroethene	0.604	0.640	-6.0	67	0.00
28 T	Bromochloromethane	0.384	0.440	-14.6	67	0.00
29 T	Tetrahydrofuran	0.089	0.105	-18.0	72	0.00
30 C	Chloroform	0.997	1.088	-9.1#	70	0.00
31 T	Cyclohexane	0.874	0.884	-1.1	67	0.00
32 T	1,1,1-Trichloroethane	0.927	1.011	-9.1	70	0.00
33 S	1,2-Dichloroethane-d4	0.512	0.551	-7.6	69	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	62	0.00
35 S	Dibromofluoromethane	0.320	0.337	-5.3	66	0.00
36 T	1,1-Dichloropropene	0.475	0.513	-8.0	68	0.00
37 T	Ethyl Acetate	0.208	0.246	-18.3	72	0.00
38 T	Carbon Tetrachloride	0.555	0.610	-9.9	69	0.00
39 T	Methylcyclohexane	0.581	0.628	-8.1	66	0.00
40 TM	Benzene	1.393	1.537	-10.3	69	0.00
41 T	Methacrylonitrile	0.120	0.142	-18.3	79	0.00
42 TM	1,2-Dichloroethane	0.386	0.439	-13.7	71	0.00
43 T	Isopropyl Acetate	0.414	0.486	-17.4	71	0.00
44 TM	Trichloroethene	0.360	0.375	-4.2	66	0.00
45 C	1,2-Dichloropropane	0.331	0.371	-12.1#	70	0.00
46 T	Dibromomethane	0.183	0.202	-10.4	69	0.00
47 T	Bromodichloromethane	0.491	0.560	-14.1	72	0.00
48 T	Methyl methacrylate	0.188	0.236	-25.5#	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021325\
 Data File : VY021196.D
 Acq On : 13 Feb 2025 17:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 00:46:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	74	0.00
50 S	Toluene-d8	1.220	1.287	-5.5	66	0.00
51 T	4-Methyl-2-Pentanone	0.208	0.256	-23.1	73	0.00
52 CM	Toluene	0.875	0.970	-10.9#	68	0.00
53 T	t-1,3-Dichloropropene	0.441	0.517	-17.2	71	0.00
54 T	cis-1,3-Dichloropropene	0.522	0.590	-13.0	69	0.00
55 T	1,1,2-Trichloroethane	0.237	0.268	-13.1	70	0.00
56 T	Ethyl methacrylate	0.317	0.375	-18.3	68	0.00
57 T	1,3-Dichloropropane	0.410	0.473	-15.4	71	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.153	-1.3	60	0.00
59 T	2-Hexanone	0.134	0.168	-25.4#	72	0.00
60 T	Dibromochloromethane	0.333	0.379	-13.8	71	0.00
61 T	1,2-Dibromoethane	0.222	0.247	-11.3	68	0.00
62 S	4-Bromofluorobenzene	0.399	0.446	-11.8	68	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	64	0.00
64 T	Tetrachloroethene	0.371	0.386	-4.0	68	0.00
65 PM	Chlorobenzene	1.113	1.168	-4.9	69	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.420	-6.1	70	0.00
67 C	Ethyl Benzene	1.969	2.120	-7.7#	69	0.00
68 T	m/p-Xylenes	0.737	0.793	-7.6	69	0.00
69 T	o-Xylene	0.687	0.741	-7.9	69	0.00
70 T	Styrene	1.145	1.269	-10.8	70	0.00
71 P	Bromoform	0.220	0.252	-14.5	74	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
73 T	Isopropylbenzene	3.906	3.888	0.5	70	0.00
74 T	N-amyl acetate	0.854	0.928	-8.7	72	0.00
75 P	1,1,2,2-Tetrachloroethane	0.650	0.653	-0.5	71	0.00
76 T	1,2,3-Trichloropropane	0.454	0.420	7.5	63	0.00
77 T	Bromobenzene	0.901	0.889	1.3	71	0.00
78 T	n-propylbenzene	4.653	4.729	-1.6	71	0.00
79 T	2-Chlorotoluene	2.665	2.626	1.5	70	0.00
80 T	1,3,5-Trimethylbenzene	3.167	3.207	-1.3	71	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.221	0.5	68	0.00
82 T	4-Chlorotoluene	2.715	2.753	-1.4	73	0.00
83 T	tert-Butylbenzene	2.878	2.901	-0.8	70	0.00
84 T	1,2,4-Trimethylbenzene	3.096	3.177	-2.6	72	0.00
85 T	sec-Butylbenzene	4.143	4.237	-2.3	72	0.00
86 T	p-Isopropyltoluene	3.442	3.556	-3.3	72	0.00
87 T	1,3-Dichlorobenzene	1.740	1.750	-0.6	72	0.00
88 T	1,4-Dichlorobenzene	1.714	1.699	0.9	70	0.00
89 T	n-Butylbenzene	3.140	3.322	-5.8	73	0.00
90 T	Hexachloroethane	0.718	0.719	-0.1	71	0.00
91 T	1,2-Dichlorobenzene	1.510	1.520	-0.7	72	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.097	0.103	-6.2	73	0.00
93 T	1,2,4-Trichlorobenzene	0.841	0.881	-4.8	69	0.00
94 T	Hexachlorobutadiene	0.557	0.572	-2.7	71	0.00
95 T	Naphthalene	1.345	1.451	-7.9	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021325\
Data File : VY021196.D
Acq On : 13 Feb 2025 17:14
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 14 00:46:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.700	0.735	-5.0	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6