

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021425\
 Data File : VY021210.D
 Acq On : 14 Feb 2025 14:38
 Operator : SY/MD
 Sample : Q1366-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-2132025

Quant Time: Feb 14 22:20:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	136771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	263641	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	244004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	94334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	81282	58.003	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.000%
35) Dibromofluoromethane	7.634	113	86815	51.382	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.760%
50) Toluene-d8	10.109	98	323834	50.344	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.408	95	99397	47.300	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.600%
Target Compounds						
16) Acetone	3.873	43	1234	5.647	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021425\
Data File : VY021210.D
Acq On : 14 Feb 2025 14:38
Operator : SY/MD
Sample : Q1366-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-2132025

Quant Time: Feb 14 22:20:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

