

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY021920\
 Data File : VY001693.D
 Acq On : 19 Feb 2020 12:03
 Operator : SY/MD
 Sample : L1362-01
 Misc : 6.87G/5ML/MSVOA Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-03-021820

Quant Time: Feb 19 13:59:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 17 12:19:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	77417	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.70	114	128946	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.49	117	103469	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.42	152	42363	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.16	65	37234	44.71	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	89.42%	
35) Dibromofluoromethane	7.74	113	35072	47.49	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	10.18	98	148032	51.16	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	12.49	95	44560	39.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.28%	
Target Compounds						
16) Acetone	3.96	43	7872	27.321	ug/l	89
20) Methylene Chloride	4.74	84	3065	2.995	ug/l	87
52) Toluene	10.24	92	3398	1.504	ug/l	98
68) m/p-Xylenes	11.70	106	1568	1.095	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY021920\
Data File : VY001693.D
Acq On : 19 Feb 2020 12:03
Operator : SY/MD
Sample : L1362-01
Misc : 6.87G/5ML/MSVOA Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-03-021820

Quant Time: Feb 19 13:59:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 17 12:19:53 2020
Response via : Initial Calibration

