

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021920\
 Data File : VY001710.D
 Acq On : 19 Feb 2020 19:26
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 20 06:42:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 17 12:19:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.487	0.503	-3.3	91	0.00
3 P	Chloromethane	0.656	0.674	-2.7	96	0.00
4 C	Vinyl Chloride	0.661	0.685	-3.6#	96	0.00
5 T	Bromomethane	0.413	0.433	-4.8	98	0.00
6 T	Chloroethane	0.407	0.427	-4.9	99	0.00
7 T	Trichlorofluoromethane	0.846	0.860	-1.7	97	0.00
8 T	Diethyl Ether	0.334	0.344	-3.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.532	0.530	0.4	94	0.00
10 T	Methyl Iodide	0.651	0.650	0.2	88	0.00
11 T	Tert butyl alcohol	0.075	0.066	12.0	105	-0.01
12 CM	1,1-Dichloroethene	0.543	0.548	-0.9#	97	0.00
13 T	Acrolein	0.054	0.069	-27.8#	127	0.00
14 T	Allyl chloride	0.966	0.982	-1.7	96	0.00
15 T	Acrylonitrile	0.171	0.178	-4.1	98	0.00
16 T	Acetone	0.186	0.153	17.7	68	0.00
17 T	Carbon Disulfide	1.811	1.746	3.6	91	0.00
18 T	Methyl Acetate	0.405	0.412	-1.7	98	0.00
19 T	Methyl tert-butyl Ether	1.510	1.577	-4.4	100	0.00
20 T	Methylene Chloride	0.661	0.623	5.7	99	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.602	2.7	94	0.00
22 T	Diisopropyl ether	1.983	2.052	-3.5	98	0.00
23 T	Vinyl Acetate	1.293	1.341	-3.7	96	0.00
24 P	1,1-Dichloroethane	1.060	1.093	-3.1	98	0.00
25 T	2-Butanone	0.253	0.240	5.1	83	0.00
26 T	2,2-Dichloropropane	0.922	0.894	3.0	95	0.00
27 T	cis-1,2-Dichloroethene	0.683	0.678	0.7	97	0.00
28 T	Bromochloromethane	0.441	0.504	-14.3	103	0.00
29 T	Tetrahydrofuran	0.153	0.159	-3.9	96	0.00
30 C	Chloroform	1.035	1.038	-0.3#	97	0.00
31 T	Cyclohexane	1.130	1.083	4.2	96	0.00
32 T	1,1,1-Trichloroethane	0.870	0.888	-2.1	97	0.00
33 S	1,2-Dichloroethane-d4	0.538	0.592	-10.0	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.286	0.291	-1.7	105	0.00
36 T	1,1-Dichloropropene	0.478	0.462	3.3	94	0.00
37 T	Ethyl Acetate	0.285	0.281	1.4	97	0.00
38 T	Carbon Tetrachloride	0.420	0.410	2.4	94	0.00
39 T	Methylcyclohexane	0.634	0.608	4.1	94	0.00
40 TM	Benzene	1.405	1.383	1.6	96	0.00
41 T	Methacrylonitrile	0.143	0.166	-16.1	104	0.00
42 TM	1,2-Dichloroethane	0.383	0.386	-0.8	98	0.00
43 T	Isopropyl Acetate	0.541	0.549	-1.5	96	0.00
44 TM	Trichloroethene	0.351	0.338	3.7	94	0.00
45 C	1,2-Dichloropropane	0.352	0.355	-0.9#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021920\
 Data File : VY001710.D
 Acq On : 19 Feb 2020 19:26
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 20 06:42:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 17 12:19:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.184	0.181	1.6	96	0.00
47 T	Bromodichloromethane	0.444	0.443	0.2	96	0.00
48 T	Methyl methacrylate	0.245	0.245	0.0	95	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	94	0.00
50 S	Toluene-d8	1.122	1.203	-7.2	103	0.00
51 T	4-Methyl-2-Pentanone	0.278	0.279	-0.4	94	0.00
52 CM	Toluene	0.876	0.868	0.9#	96	0.00
53 T	t-1,3-Dichloropropene	0.491	0.497	-1.2	95	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.562	0.5	96	0.00
55 T	1,1,2-Trichloroethane	0.271	0.273	-0.7	97	0.00
56 T	Ethyl methacrylate	0.407	0.421	-3.4	97	0.00
57 T	1,3-Dichloropropane	0.482	0.486	-0.8	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.170	0.146	14.1	100	0.00
59 T	2-Hexanone	0.209	0.198	5.3	85	0.00
60 T	Dibromochloromethane	0.299	0.300	-0.3	94	0.00
61 T	1,2-Dibromoethane	0.257	0.264	-2.7	98	0.00
62 S	4-Bromofluorobenzene	0.436	0.437	-0.2	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.314	0.297	5.4	94	0.00
65 PM	Chlorobenzene	0.998	0.961	3.7	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.337	0.333	1.2	96	0.00
67 C	Ethyl Benzene	1.853	1.806	2.5#	94	0.00
68 T	m/p-Xylenes	0.692	0.675	2.5	94	0.00
69 T	o-Xylene	0.658	0.637	3.2	93	0.00
70 T	Styrene	1.143	1.116	2.4	93	0.00
71 P	Bromoform	0.192	0.186	3.1	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.809	3.789	0.5	94	0.00
74 T	N-amyl acetate	1.216	1.208	0.7	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.813	0.831	-2.2	92	0.00
76 T	1,2,3-Trichloropropane	0.574	0.587	-2.3	105	0.00
77 T	Bromobenzene	0.826	0.807	2.3	92	0.00
78 T	n-propylbenzene	4.639	4.645	-0.1	93	0.00
79 T	2-Chlorotoluene	2.602	2.599	0.1	94	0.00
80 T	1,3,5-Trimethylbenzene	3.153	3.122	1.0	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.302	0.300	0.7	86	0.00
82 T	4-Chlorotoluene	2.744	2.706	1.4	90	0.00
83 T	tert-Butylbenzene	2.637	2.648	-0.4	93	0.00
84 T	1,2,4-Trimethylbenzene	3.144	3.127	0.5	92	0.00
85 T	sec-Butylbenzene	3.879	3.829	1.3	91	0.00
86 T	p-Isopropyltoluene	3.369	3.302	2.0	90	0.00
87 T	1,3-Dichlorobenzene	1.601	1.535	4.1	90	0.00
88 T	1,4-Dichlorobenzene	1.619	1.534	5.3	89	0.00
89 T	n-Butylbenzene	3.443	3.308	3.9	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021920\
Data File : VY001710.D
Acq On : 19 Feb 2020 19:26
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 20 06:42:40 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 17 12:19:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.635	0.619	2.5	89	0.00
91 T	1,2-Dichlorobenzene	1.478	1.415	4.3	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.134	0.135	-0.7	92	0.00
93 T	1,2,4-Trichlorobenzene	0.979	0.880	10.1	84	0.00
94 T	Hexachlorobutadiene	0.487	0.427	12.3	85	0.00
95 T	Naphthalene	2.258	2.147	4.9	87	0.00
96 T	1,2,3-Trichlorobenzene	0.862	0.777	9.9	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6