


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv021920\  
 Data File : VY001688.D  
 Acq On : 19 Feb 2020 09:36  
 Operator : SY/MD  
 Sample : BFB  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y021720S.M  
 Title : SW846 8260  
 Last Update : Mon Feb 17 12:19:53 2020



AutoFind: Scans 1763, 1764, 1765; Background Corrected with Scan 1755

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 19.6      | 23394   | PASS             |
| 75          | 95           | 30           | 60           | 51.5      | 61354   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 119133  | PASS             |
| 96          | 95           | 5            | 9            | 6.6       | 7872    | PASS             |
| 173         | 174          | 0.00         | 2            | 1.1       | 950     | PASS             |
| 174         | 95           | 50           | 100          | 72.8      | 86746   | PASS             |
| 175         | 174          | 5            | 9            | 7.7       | 6648    | PASS             |
| 176         | 174          | 95           | 101          | 96.5      | 83701   | PASS             |
| 177         | 176          | 5            | 9            | 6.7       | 5590    | PASS             |