

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY022020\
 Data File : VY001722.D
 Acq On : 20 Feb 2020 14:28
 Operator : SY/MD
 Sample : L1589-02
 Misc : 7.67G/5ML/MSVOA Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB4-4.5-5.0

Quant Time: Feb 21 08:31:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 17 12:19:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.82	168	210193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.71	114	387017	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	350691	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.44	152	165149	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.17	65	120408	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
35) Dibromofluoromethane	7.75	113	113134	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
50) Toluene-d8	10.19	98	459279	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
62) 4-Bromofluorobenzene	12.49	95	168612	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
Target Compounds						
52) Toluene	10.25	92	17958	2.648	ug/l	95
67) Ethyl Benzene	11.61	91	55503	4.272	ug/l	96
68) m/p-Xylenes	11.71	106	88096	18.151	ug/l	97
69) o-Xylene	12.04	106	28708	6.222	ug/l	98
78) n-propylbenzene	12.68	91	31265	2.040	ug/l	96
80) 1,3,5-Trimethylbenzene	12.82	105	32690	3.139	ug/l	93
84) 1,2,4-Trimethylbenzene	13.13	105	139196	13.406	ug/l	100
95) Naphthalene	15.24	128	101580	13.619	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY022020\
Data File : VY001722.D
Acq On : 20 Feb 2020 14:28
Operator : SY/MD
Sample : L1589-02
Misc : 7.67G/5ML/MSVOA Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB4-4.5-5.0

Quant Time: Feb 21 08:31:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 17 12:19:53 2020
Response via : Initial Calibration

