

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY022020\
 Data File : VY001725.D
 Acq On : 20 Feb 2020 16:04
 Operator : SY/MD
 Sample : L1624-02
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WASTE-2

Quant Time: Feb 21 08:37:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 17 12:19:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.82	168	22831	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.71	114	38061	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	27773	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.43	152	10288	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.16	65	9966	40.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.16%	
35) Dibromofluoromethane	7.74	113	9902	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
50) Toluene-d8	10.19	98	42080	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.49	95	11725	35.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.68%	
Target Compounds						
52) Toluene	10.25	92	2108	3.161	ug/l	93
67) Ethyl Benzene	11.60	91	6538	6.354	ug/l	95
68) m/p-Xylenes	11.70	106	9389	24.427	ug/l	97
69) o-Xylene	12.03	106	3774	10.328	ug/l	98
78) n-propylbenzene	12.67	91	4141	4.338	ug/l	99
80) 1,3,5-Trimethylbenzene	12.81	105	5136	7.916	ug/l	99
84) 1,2,4-Trimethylbenzene	13.13	105	18838	29.125	ug/l	95
95) Naphthalene	15.25	128	7042	15.156	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY022020\
Data File : VY001725.D
Acq On : 20 Feb 2020 16:04
Operator : SY/MD
Sample : L1624-02
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WASTE-2

Quant Time: Feb 21 08:37:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y021720S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 17 12:19:53 2020
Response via : Initial Calibration

