

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022221\
 Data File : VY004036.D
 Acq On : 22 Feb 2021 15:38
 Operator : SY/MD
 Sample : VY0222SBL01
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0222SBL01

Quant Time: Feb 23 03:44:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 23 03:42:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	63331	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	98610	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	89916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	44339	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	41456	58.66	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.32%
35) Dibromofluoromethane	7.722	113	32931	55.39	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.78%
50) Toluene-d8	10.179	98	130775	54.86	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.72%
62) 4-Bromofluorobenzene	12.477	95	45110	53.57	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.14%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022221\
Data File : VY004036.D
Acq On : 22 Feb 2021 15:38
Operator : SY/MD
Sample : VY0222SBL01
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0222SBL01

Quant Time: Feb 23 03:44:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022221S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 23 03:42:58 2021
Response via : Initial Calibration

