

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022224\
 Data File : VY017456.D
 Acq On : 22 Feb 2024 14:26
 Operator : SY/MD
 Sample : P1560-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR20010

Quant Time: Feb 23 00:30:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 20 06:16:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	125622	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.691	114	232929	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	204264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	47675	44.424	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.840%
35) Dibromofluoromethane	7.716	113	42935	31.489	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	62.980%
50) Toluene-d8	10.185	98	199462	37.719	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	75.440%
62) 4-Bromofluorobenzene	12.483	95	53216	30.859	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	61.720%
Target Compounds						
20) Methylene Chloride	4.716	84	20561	9.487	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022224\
Data File : VY017456.D
Acq On : 22 Feb 2024 14:26
Operator : SY/MD
Sample : P1560-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR20010

Quant Time: Feb 23 00:30:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021924S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 20 06:16:24 2024
Response via : Initial Calibration

