

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022224\
 Data File : VY017460.D
 Acq On : 22 Feb 2024 16:01
 Operator : SY/MD
 Sample : P1537-12
 Misc : 5.33g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 12

Quant Time: Feb 23 00:31:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 20 06:16:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	135814	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.691	114	253768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	214347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	69200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	48378	41.696	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.400%
35) Dibromofluoromethane	7.716	113	60154	40.495	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.000%
50) Toluene-d8	10.185	98	212886	36.952	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	73.900%
62) 4-Bromofluorobenzene	12.489	95	52095	27.904	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	55.800%
Target Compounds						
20) Methylene Chloride	4.716	84	23944	10.751	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022224\
Data File : VY017460.D
Acq On : 22 Feb 2024 16:01
Operator : SY/MD
Sample : P1537-12
Misc : 5.33g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
12

Quant Time: Feb 23 00:31:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021924S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 20 06:16:24 2024
Response via : Initial Calibration

