

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022421\
 Data File : VY004044.D
 Acq On : 24 Feb 2021 10:16
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 01:11:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 23 03:42:58 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.475	0.471	0.8	87	0.00
3 P	Chloromethane	0.709	0.754	-6.3	91	0.00
4 C	Vinyl Chloride	0.687	0.748	-8.9#	90	0.00
5 T	Bromomethane	0.461	0.496	-7.6	90	0.00
6 T	Chloroethane	0.415	0.468	-12.8	89	0.00
7 T	Trichlorofluoromethane	0.860	0.958	-11.4	88	0.00
8 T	Diethyl Ether	0.275	0.295	-7.3	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.489	-6.3	80	0.00
10 T	Methyl Iodide	0.577	0.632	-9.5	77	0.00
11 T	Tert butyl alcohol	0.083	0.052	37.3#	69	0.00
12 CM	1,1-Dichloroethene	0.462	0.486	-5.2#	79	0.00
13 T	Acrolein	0.055	0.056	-1.8	75	0.00
14 T	Allyl chloride	0.875	0.899	-2.7	77	0.00
15 T	Acrylonitrile	0.145	0.146	-0.7	78	0.00
16 T	Acetone	0.141	0.168	-19.1	94	0.00
17 T	Carbon Disulfide	1.491	1.503	-0.8	75	0.00
18 T	Methyl Acetate	0.331	0.327	1.2	77	0.00
19 T	Methyl tert-butyl Ether	1.355	1.402	-3.5	80	0.00
20 T	Methylene Chloride	0.570	0.538	5.6	77	0.00
21 T	trans-1,2-Dichloroethene	0.519	0.535	-3.1	77	0.00
22 T	Diisopropyl ether	1.720	1.762	-2.4	77	0.00
23 T	Vinyl Acetate	1.165	1.196	-2.7	79	0.00
24 P	1,1-Dichloroethane	0.920	0.949	-3.2	77	0.00
25 T	2-Butanone	0.201	0.219	-9.0	86	0.00
26 T	2,2-Dichloropropane	0.851	0.918	-7.9	83	0.00
27 T	cis-1,2-Dichloroethene	0.570	0.597	-4.7	79	0.00
28 T	Bromochloromethane	0.406	0.388	4.4	73	0.00
29 T	Tetrahydrofuran	0.130	0.125	3.8	76	0.00
30 C	Chloroform	0.933	0.992	-6.3#	80	0.00
31 T	Cyclohexane	0.936	0.916	2.1	77	0.00
32 T	1,1,1-Trichloroethane	0.857	0.912	-6.4	81	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.573	-2.7	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.301	0.310	-3.0	74	0.00
36 T	1,1-Dichloropropene	0.501	0.512	-2.2	78	0.00
37 T	Ethyl Acetate	0.296	0.290	2.0	78	0.00
38 T	Carbon Tetrachloride	0.517	0.549	-6.2	81	0.00
39 T	Methylcyclohexane	0.605	0.631	-4.3	78	0.00
40 TM	Benzene	1.391	1.427	-2.6	78	0.00
41 T	Methacrylonitrile	0.156	0.149	4.5	69	0.00
42 TM	1,2-Dichloroethane	0.457	0.481	-5.3	81	0.00
43 T	Isopropyl Acetate	0.558	0.568	-1.8	78	0.00
44 TM	Trichloroethene	0.384	0.399	-3.9	78	0.00
45 C	1,2-Dichloropropane	0.341	0.358	-5.0#	78	0.00
46 T	Dibromomethane	0.191	0.197	-3.1	78	0.00
47 T	Bromodichloromethane	0.487	0.516	-6.0	80	0.00
48 T	Methyl methacrylate	0.273	0.276	-1.1	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022421\
 Data File : VY004044.D
 Acq On : 24 Feb 2021 10:16
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 01:11:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 23 03:42:58 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	75	0.00
50 S	Toluene-d8	1.209	1.252	-3.6	74	0.00
51 T	4-Methyl-2-Pentanone	0.288	0.294	-2.1	79	0.00
52 CM	Toluene	0.870	0.921	-5.9#	80	0.00
53 T	t-1,3-Dichloropropene	0.524	0.552	-5.3	79	0.00
54 T	cis-1,3-Dichloropropene	0.587	0.618	-5.3	79	0.00
55 T	1,1,2-Trichloroethane	0.270	0.281	-4.1	79	0.00
56 T	Ethyl methacrylate	0.386	0.408	-5.7	78	0.00
57 T	1,3-Dichloropropane	0.478	0.501	-4.8	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.175	0.157	10.3	78	0.00
59 T	2-Hexanone	0.203	0.214	-5.4	81	0.00
60 T	Dibromochloromethane	0.341	0.367	-7.6	81	0.00
61 T	1,2-Dibromoethane	0.265	0.276	-4.2	80	0.00
62 S	4-Bromofluorobenzene	0.427	0.435	-1.9	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.461	0.484	-5.0	79	0.00
65 PM	Chlorobenzene	1.007	1.058	-5.1	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.410	-6.5	80	0.00
67 C	Ethyl Benzene	1.889	1.992	-5.5#	79	0.00
68 T	m/p-Xylenes	0.697	0.740	-6.2	79	0.00
69 T	o-Xylene	0.655	0.694	-6.0	79	0.00
70 T	Styrene	1.125	1.195	-6.2	79	0.00
71 P	Bromoform	0.246	0.259	-5.3	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
73 T	Isopropylbenzene	3.550	3.770	-6.2	80	0.00
74 T	N-amyl acetate	1.072	1.130	-5.4	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.646	0.652	-0.9	76	0.00
76 T	1,2,3-Trichloropropane	0.502	0.516	-2.8	79	0.00
77 T	Bromobenzene	0.874	0.925	-5.8	79	0.00
78 T	n-propylbenzene	4.250	4.530	-6.6	79	0.00
79 T	2-Chlorotoluene	2.411	2.526	-4.8	79	0.00
80 T	1,3,5-Trimethylbenzene	3.028	3.232	-6.7	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.270	0.270	0.0	74	0.00
82 T	4-Chlorotoluene	2.530	2.682	-6.0	79	0.00
83 T	tert-Butylbenzene	2.593	2.781	-7.3	80	0.00
84 T	1,2,4-Trimethylbenzene	3.034	3.237	-6.7	79	0.00
85 T	sec-Butylbenzene	3.552	3.801	-7.0	78	0.00
86 T	p-Isopropyltoluene	3.308	3.596	-8.7	80	0.00
87 T	1,3-Dichlorobenzene	1.664	1.759	-5.7	79	0.00
88 T	1,4-Dichlorobenzene	1.653	1.744	-5.5	78	0.00
89 T	n-Butylbenzene	3.072	3.369	-9.7	79	0.00
90 T	Hexachloroethane	0.590	0.631	-6.9	79	0.00
91 T	1,2-Dichlorobenzene	1.472	1.534	-4.2	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.124	0.129	-4.0	79	0.00
93 T	1,2,4-Trichlorobenzene	1.058	1.119	-5.8	79	0.00
94 T	Hexachlorobutadiene	0.668	0.716	-7.2	79	0.00
95 T	Naphthalene	1.928	1.978	-2.6	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022421\
Data File : VY004044.D
Acq On : 24 Feb 2021 10:16
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 25 01:11:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022221S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 23 03:42:58 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.902	0.936	-3.8	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6