

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022422\
 Data File : VY007624.D
 Acq On : 24 Feb 2022 17:50
 Operator : SY/MD
 Sample : N1584-04
 Misc : 5.94g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D11-12.5

Quant Time: Feb 25 05:43:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 24 07:13:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	90851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	155183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	151286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64695	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	52755	52.198	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 104.400%		
35) Dibromofluoromethane	7.722	113	50425	51.497	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 103.000%		
50) Toluene-d8	10.179	98	180669	49.880	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 99.760%		
62) 4-Bromofluorobenzene	12.483	95	63613	51.378	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 102.760%		
Target Compounds						
						Qvalue
3) Chloromethane	2.113	50	786	1.451	ug/l #	74
16) Acetone	3.936	43	3119	13.551	ug/l	97
17) Carbon Disulfide	4.204	76	13208	5.179	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022422\
Data File : VY007624.D
Acq On : 24 Feb 2022 17:50
Operator : SY/MD
Sample : N1584-04
Misc : 5.94g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D11-12.5

Quant Time: Feb 25 05:43:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022322S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 24 07:13:58 2022
Response via : Initial Calibration

