

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022422\
 Data File : VY007630.D
 Acq On : 24 Feb 2022 20:11
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 25 05:44:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 24 07:13:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	81567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	140001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	139684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	68236	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	44340	48.865	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.740%
35) Dibromofluoromethane	7.716	113	44236	50.075	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.160%
50) Toluene-d8	10.179	98	162208	49.658	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.320%
62) 4-Bromofluorobenzene	12.483	95	59531	53.295	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022422\
Data File : VY007630.D
Acq On : 24 Feb 2022 20:11
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 25 05:44:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022322S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 24 07:13:58 2022
Response via : Initial Calibration

