



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022425\
Data File : VY021299.D
Acq On : 24 Feb 2025 12:44
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022425S.M
Title : SW846 8260
Last Update : Tue Feb 25 01:06:48 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.1	22171	PASS
75	95	30	60	57.2	52723	PASS
95	95	100	100	100.0	92187	PASS
96	95	5	9	6.8	6248	PASS
173	174	0.00	2	0.9	702	PASS
174	95	50	100	83.1	76568	PASS
175	174	5	9	7.8	5952	PASS
176	174	95	101	95.4	73053	PASS
177	176	5	9	6.7	4926	PASS