

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022625\
 Data File : VY021328.D
 Acq On : 26 Feb 2025 13:54
 Operator : SY/MD
 Sample : Q1401-07
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-VPB-192-GW-825-827

Quant Time: Feb 27 00:37:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 26 02:09:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	237643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	467253	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	479781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	211227	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	165185	61.274	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.540%
35) Dibromofluoromethane	7.634	113	159219	51.975	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.960%
50) Toluene-d8	10.103	98	584543	49.374	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.401	95	222969	55.961	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022625\
Data File : VY021328.D
Acq On : 26 Feb 2025 13:54
Operator : SY/MD
Sample : Q1401-07
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-VPB-192-GW-825-827

Quant Time: Feb 27 00:37:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 26 02:09:13 2025
Response via : Initial Calibration

