

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022625\
 Data File : VY021329.D
 Acq On : 26 Feb 2025 14:17
 Operator : SY/MD
 Sample : Q1401-08
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-VPB-192-GW-860-862

Quant Time: Feb 27 00:37:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 26 02:09:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	234926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	466686	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	492462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	224351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	180680	67.797	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.600%
35) Dibromofluoromethane	7.634	113	162833	53.220	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.440%
50) Toluene-d8	10.103	98	586525	49.602	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.402	95	232790	58.497	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	117.000%
Target Compounds						
18) Methyl Acetate	4.391	43	1330	1.065	ug/l	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022625\
Data File : VY021329.D
Acq On : 26 Feb 2025 14:17
Operator : SY/MD
Sample : Q1401-08
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-VPB-192-GW-860-862

Quant Time: Feb 27 00:37:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 26 02:09:13 2025
Response via : Initial Calibration

