

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022723\
 Data File : VY012739.D
 Acq On : 27 Feb 2023 17:14
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 00:23:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.514	0.443	13.8	81	0.00
3 P	Chloromethane	0.687	0.506	26.3#	72	0.00
4 C	Vinyl Chloride	0.568	0.492	13.4#	79	0.00
5 T	Bromomethane	0.311	0.303	2.6	91	0.00
6 T	Chloroethane	0.362	0.303	16.3	81	0.00
7 T	Trichlorofluoromethane	0.916	0.850	7.2	85	0.00
8 T	Diethyl Ether	0.328	0.305	7.0	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.527	3.7	89	0.00
10 T	Methyl Iodide	0.716	0.677	5.4	84	0.00
11 T	Tert butyl alcohol	0.065	0.054	16.9	81	0.00
12 CM	1,1-Dichloroethene	0.528	0.493	6.6#	86	0.00
13 T	Acrolein	0.082	0.067	18.3	66	0.00
14 T	Allyl chloride	1.025	0.920	10.2	79	0.00
15 T	Acrylonitrile	0.174	0.162	6.9	76	0.00
16 T	Acetone	0.193	0.183	5.2	77	0.00
17 T	Carbon Disulfide	1.778	1.478	16.9	76	0.00
18 T	Methyl Acetate	0.621	0.561	9.7	75	0.00
19 T	Methyl tert-butyl Ether	1.421	1.348	5.1	81	0.00
20 T	Methylene Chloride	0.734	0.630	14.2	90	0.00
21 T	trans-1,2-Dichloroethene	0.575	0.542	5.7	85	0.00
22 T	Diisopropyl ether	1.913	1.791	6.4	82	0.00
23 T	Vinyl Acetate	1.185	1.096	7.5	77	0.00
24 P	1,1-Dichloroethane	1.061	0.984	7.3	84	0.00
25 T	2-Butanone	0.256	0.231	9.8	73	0.00
26 T	2,2-Dichloropropane	0.925	0.862	6.8	88	0.00
27 T	cis-1,2-Dichloroethene	0.629	0.612	2.7	88	0.00
28 T	Bromochloromethane	0.425	0.404	4.9	80	0.00
29 T	Tetrahydrofuran	0.156	0.143	8.3	73	0.00
30 C	Chloroform	1.028	1.001	2.6#	88	0.00
31 T	Cyclohexane	1.026	0.862	16.0	81	0.00
32 T	1,1,1-Trichloroethane	0.930	0.889	4.4	88	0.00
33 S	1,2-Dichloroethane-d4	0.591	0.528	10.7	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.338	0.327	3.3	85	0.00
36 T	1,1-Dichloropropene	0.535	0.488	8.8	84	0.00
37 T	Ethyl Acetate	0.361	0.309	14.4	69	0.00
38 T	Carbon Tetrachloride	0.554	0.531	4.2	88	0.00
39 T	Methylcyclohexane	0.618	0.592	4.2	85	0.00
40 TM	Benzene	1.494	1.443	3.4	87	0.00
41 T	Methacrylonitrile	0.176	0.171	2.8	90	0.00
42 TM	1,2-Dichloroethane	0.463	0.436	5.8	83	0.00
43 T	Isopropyl Acetate	0.624	0.568	9.0	75	0.00
44 TM	Trichloroethene	0.386	0.386	0.0	91	0.00
45 C	1,2-Dichloropropane	0.384	0.373	2.9#	86	0.00
46 T	Dibromomethane	0.212	0.211	0.5	86	0.00
47 T	Bromodichloromethane	0.511	0.509	0.4	90	0.00
48 T	Methyl methacrylate	0.291	0.268	7.9	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022723\
 Data File : VY012739.D
 Acq On : 27 Feb 2023 17:14
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 00:23:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	77	0.00
50 S	Toluene-d8	1.265	1.225	3.2	85	0.00
51 T	4-Methyl-2-Pentanone	0.337	0.315	6.5	75	0.00
52 CM	Toluene	0.885	0.900	-1.7#	90	0.00
53 T	t-1,3-Dichloropropene	0.534	0.527	1.3	86	0.00
54 T	cis-1,3-Dichloropropene	0.604	0.603	0.2	87	0.00
55 T	1,1,2-Trichloroethane	0.288	0.297	-3.1	91	0.00
56 T	Ethyl methacrylate	0.394	0.404	-2.5	83	0.00
57 T	1,3-Dichloropropane	0.507	0.507	0.0	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.207	-5.6	82	0.00
59 T	2-Hexanone	0.239	0.229	4.2	74	0.00
60 T	Dibromochloromethane	0.349	0.365	-4.6	92	0.00
61 T	1,2-Dibromoethane	0.276	0.283	-2.5	89	0.00
62 S	4-Bromofluorobenzene	0.405	0.396	2.2	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.364	0.353	3.0	91	0.00
65 PM	Chlorobenzene	1.072	1.064	0.7	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.404	-1.8	93	0.00
67 C	Ethyl Benzene	1.917	1.918	-0.1#	92	0.00
68 T	m/p-Xylenes	0.731	0.744	-1.8	93	0.00
69 T	o-Xylene	0.677	0.694	-2.5	92	0.00
70 T	Styrene	1.135	1.181	-4.1	93	0.00
71 P	Bromoform	0.254	0.258	-1.6	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.695	3.682	0.4	94	0.00
74 T	N-amyl acetate	1.215	1.103	9.2	77	0.00
75 P	1,1,2,2-Tetrachloroethane	0.831	0.812	2.3	89	0.00
76 T	1,2,3-Trichloropropane	0.623	0.596	4.3	88	0.00
77 T	Bromobenzene	0.885	0.880	0.6	94	0.00
78 T	n-propylbenzene	4.622	4.546	1.6	94	0.00
79 T	2-Chlorotoluene	2.568	2.486	3.2	93	0.00
80 T	1,3,5-Trimethylbenzene	3.086	3.035	1.7	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.294	0.286	2.7	87	0.00
82 T	4-Chlorotoluene	2.663	2.608	2.1	94	0.00
83 T	tert-Butylbenzene	2.636	2.633	0.1	94	0.00
84 T	1,2,4-Trimethylbenzene	3.017	2.967	1.7	92	0.00
85 T	sec-Butylbenzene	4.010	3.989	0.5	95	0.00
86 T	p-Isopropyltoluene	3.303	3.294	0.3	95	0.00
87 T	1,3-Dichlorobenzene	1.734	1.701	1.9	94	0.00
88 T	1,4-Dichlorobenzene	1.758	1.699	3.4	94	0.00
89 T	n-Butylbenzene	3.163	3.119	1.4	94	0.00
90 T	Hexachloroethane	0.670	0.650	3.0	96	0.00
91 T	1,2-Dichlorobenzene	1.549	1.537	0.8	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.137	0.125	8.8	83	0.00
93 T	1,2,4-Trichlorobenzene	0.979	0.897	8.4	88	0.00
94 T	Hexachlorobutadiene	0.591	0.523	11.5	89	0.00
95 T	Naphthalene	1.979	1.829	7.6	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022723\
Data File : VY012739.D
Acq On : 27 Feb 2023 17:14
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 28 00:23:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.867	0.781	9.9	86	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6