

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
 Data File : VY017497.D
 Acq On : 27 Feb 2024 15:30
 Operator : SY/MD
 Sample : P1564-19RE
 Misc : 3.52g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-18RE

Quant Time: Feb 28 00:17:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	204381	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	179566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	74889	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	77429	71.994	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.980%
35) Dibromofluoromethane	7.716	113	82339	57.815	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.620%
50) Toluene-d8	10.179	98	294036	59.049	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	118.100%
62) 4-Bromofluorobenzene	12.483	95	84571	54.042	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
Data File : VY017497.D
Acq On : 27 Feb 2024 15:30
Operator : SY/MD
Sample : P1564-19RE
Misc : 3.52g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-18RE

Quant Time: Feb 28 00:17:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

