

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
 Data File : VY017502.D
 Acq On : 27 Feb 2024 17:27
 Operator : SY/MD
 Sample : P1564-11
 Misc : 3.55g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-14

Quant Time: Feb 28 00:19:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	115235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	211618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	194284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	79990	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	66626	61.006	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.020%
35) Dibromofluoromethane	7.716	113	72382	49.085	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.180%
50) Toluene-d8	10.179	98	240841	46.712	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.420%
62) 4-Bromofluorobenzene	12.483	95	68865	42.500	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
Data File : VY017502.D
Acq On : 27 Feb 2024 17:27
Operator : SY/MD
Sample : P1564-11
Misc : 3.55g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-14

Quant Time: Feb 28 00:19:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

