

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
 Data File : VY017503.D
 Acq On : 27 Feb 2024 17:50
 Operator : SY/MD
 Sample : P1564-07RE
 Misc : 3.84g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-12RE

Quant Time: Feb 28 00:19:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	107761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	204663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	188201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	78067	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	55576	54.418	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.840%
35) Dibromofluoromethane	7.722	113	58558	41.060	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.120%
50) Toluene-d8	10.179	98	191563	38.417	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	76.840%
62) 4-Bromofluorobenzene	12.483	95	55775	35.592	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	71.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
Data File : VY017503.D
Acq On : 27 Feb 2024 17:50
Operator : SY/MD
Sample : P1564-07RE
Misc : 3.84g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-12RE

Quant Time: Feb 28 00:19:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

