

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
 Data File : VY017506.D
 Acq On : 27 Feb 2024 18:59
 Operator : SY/MD
 Sample : P1564-01
 Misc : 3.33g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-09

Quant Time: Feb 28 00:21:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	108449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	202526	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	183460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75403	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	64532	62.786	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.580%
35) Dibromofluoromethane	7.722	113	70811	50.176	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.360%
50) Toluene-d8	10.185	98	236301	47.889	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.780%
62) 4-Bromofluorobenzene	12.483	95	66833	43.098	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
Data File : VY017506.D
Acq On : 27 Feb 2024 18:59
Operator : SY/MD
Sample : P1564-01
Misc : 3.33g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-09

Quant Time: Feb 28 00:21:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

