



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022724\
Data File : VY017488.D
Acq On : 27 Feb 2024 09:40
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Title : SW846 8260
Last Update : Sat Feb 24 04:41:20 2024



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1742

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	5788	PASS
75	95	30	60	49.0	16597	PASS
95	95	100	100	100.0	33859	PASS
96	95	5	9	6.6	2247	PASS
173	174	0.00	2	1.5	477	PASS
174	95	50	100	96.0	32499	PASS
175	174	5	9	7.7	2506	PASS
176	174	95	101	96.9	31493	PASS
177	176	5	9	6.3	1986	PASS